

# **EU DECLARATION OF CONFORMITY FOR ATTACHMENT OF 'CE' MARK**



We, the manufacturer: Ecolab Limited.  
52 Royce Close,  
West Portway  
Andover, UK SP10 3TS  
Tel: +44 (0) 1264 - 835835

declare under our sole responsibility that the appliance:

## **Bioquell SeQure (FS-1) Suite**

Serial Number as stated on the unit's rating plate

complies with the Essential Requirements of the Machinery Directive 2006/42/EC, Radio Equipment Directive 2014/53/EU, Restriction of Hazardous Substances in Electrical and Electronic Equipment 2011/65/EU, and the Battery Regulation 2023/1542 by means of the following harmonised standards:

Machinery Directive: by means of BS EN 12100:2010

Radio Equipment Directive by means of:

Article 3.1a	RF Exposure: BS EN 62311
Article 3.1a	BS EN 61010-1:2010+A1:2019
Article 3.1b	BS EN 61326-2-1:2013 using ETSI EN 301 489-3 V2.1.1, & ETSI EN 301 489-17 V3.1.1 (ETSI EN 301 489-1 V2.1.1)
Article 3.2	EN 300 328 v 2.1.1 EN 300 330 v 2.1.1

Restriction of Hazardous Substances in Electrical and Electronic Equipment by means of BS EN IEC 63000:2018.

A copy of the Technical File for this equipment is available from:

Ecolab Regulatory Affairs – Bioquell,  
Avda. Baix Llobregat 3, 1<sup>a</sup>,  
08970 SANT JOAN DESPI, España

Ecolab Limited, 52 Royce Close, West  
Portway, Andover, SP10 3TS, UK

*Authorised signatory of manufacturer:*

**O.R.Cumberlege, Bioquell R&D Group Leader**

7<sup>th</sup> November 2024